



## Current Probe *(Preliminary-subject to change)*

Part No:	Freq	Zt	Max A 400Hz	Max CW	Max Pulse	Window
OCP212	1MHz-1GHz	1-7	200	20	50	1.25
OCP217	20Hz-150MHz	1	450	4.2	100	1.25
OCP218	10kHz-500MHz	1-5	200	40	60	1.25
OCP219	10kHz-150MHz	1	350	4.2	100	1.25
OCP220	10kHz-100MHz	5	350	42	100	1.25
OPC405	10kHz-400MHz	8				2.00
OCP321	100kHz-500MHz	8	200	25	50	2.62
OCP324	10KHz-300MHz	8	300	62	62	2.62



## BCI Probes *(Preliminary-subject to change)*

Part No	Frequency	Wind A.	Watts	Window
OIP257	100kHz-1GHz		100	1.5
OIP400	10KHz-400MHz		100	2



## Calibration Fixtures

Part No.	Frequency	Width	Height	Center
OFC120	20Hz-500MHz	10	7	6
OFC121	100kHz-1GHz	10	7	7
OFC403	10kHz-400MHz	12	7	7

